



***DeepChek<sup>®</sup> Assay***  
***NS3 Genotyping and Drug Resistance (RUO)***  
**V1**  
**User Guide**



Version 1 – Revision 0

**For Research Use Only (RUO). Not for use in diagnostic procedures. No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of disease.**

**REF** 108A24 (old reference: K-16-NS3DR)

## Contents

Application	3
Principles of the assay	3
Assay components	4
Reagent storage and handling	5
Materials required but not provided	5
Warnings and precautions	5
Starting	6
RNA Extraction	6
RT-PCR Workflow Overview	7
RT and PCR Step-by-Step Workflow for NS3 Target	8
Nested PCR Step-by-Step Workflow for NS3 Target	9
RT-PCR Troubleshooting Guide	10
PCR Products Purification	10
Sequencing	10
1. Purification	10
2. Sanger sequencing	10
3. NGS	12
Data Analysis	12
1. Sanger	12
2. NGS	13
Product quality control	13
Symbols	13
Contact Information	13
Manufacturer	13

## Application

For Research Use Only (RUO). Not for use in diagnostic procedures. No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of disease.

The DeepChek® Assay NS3 genotyping and Drug Resistance (DR) (RUO) kit is a single tube system which utilizes PCR technology for amplifying relevant portions of the human C virus (HCV) NS3 gene from input RNA extracted from Plasma/serum.

This nucleic acid amplification method might aid in the typing of HCV viruses and DR. This test is NOT intended to be used as a screening or confirmation test for the detection, confirmation and quantification of HCV infection.

The DeepChek® Assay NS3 genotyping and Drug Resistance (RUO) is intended for use by trained laboratory personnel specifically instructed and trained in the techniques of PCR, SANGER and next generation sequencing (NGS) workflow.

## Principles of the assay

The DeepChek® Assay NS3 genotyping and Drug Resistance (RUO) is a reverse transcriptase-polymerase chain reaction test which includes primers, reverse and forward, designed to amplify HCV RNA from plasma/serum specimens.

First, reverse transcription is performed and the HCV reverse primers anneal to their respective targets and are extended during a prolonged incubation period. After a denaturation step, in which the temperature of the reaction is raised above the melting point of the double-stranded DNA, a second primer anneals to the DNA strand and is extended by the DNA polymerase activity of the enzyme to create a double-stranded DNA product.

During each round of thermal cycling, amplification products dissociate to single strands at high temperature allowing primer annealing and extension as the temperature is lowered. Exponential amplification of the product is achieved through repeated cycling between high and low temperatures, resulting in a billion-fold or greater amplification of target sequences.

DeepChek® Assay NS3 genotyping and Drug Resistance (RUO) is performed on a PCR instrument.

Subsequently, the amplicons can be used for SANGER or next generation sequencing and analyzed with a downstream analysis software to list in a report HCV genotypes according to available public reference knowledge databases.

## Assay components

The *DeepChek® Assay NS3 Genotyping and Drug Resistance V1 (RUO)* is provided in one model of 24 reactions (REF 108A24 / OLD REF K-16-NS3DR V1).

Label	Volume for 24 Rxn	Color cap	Storage
<b>RT-PCR</b>			
RT-PCR Buffer 2X	480 µL	Green	-25°C to - 15 °C
RP Solution	10 µL	Purple	-25°C to - 15 °C
RT Enzyme	10 µL	White	-25°C to - 15 °C
BP Solution	5 µL	Pink	-25°C to - 15 °C
PCR Enzyme	15 µL	White	-25°C to - 15 °C
H <sub>2</sub> O	500 µL	Blue	-25°C to - 15 °C
NS3 FOR <sub>1</sub> PCR Primers	20 µL	Yellow	-25°C to - 15 °C
NS3 FOR <sub>2</sub> PCR Primers	20 µL	Yellow	-25°C to - 15 °C
NS3 REV PCR Primers	40 µL	Yellow	-25°C to - 15 °C
<b>Nested PCR</b>			
Nested PCR Buffer 2X	400 µL	Green	-25°C to - 15 °C
Nested PCR Enzyme	10 µL	White	-25°C to - 15 °C
H <sub>2</sub> O	500 µL	Blue	-25°C to - 15 °C
NS3 FOR <sub>1</sub> Nested PCR Primers	20 µL	Yellow	-25°C to - 15 °C
NS3 FOR <sub>2</sub> Nested PCR Primers	20 µL	Yellow	-25°C to - 15 °C
NS3 REV Nested PCR Primers	35 µL	Yellow	-25°C to - 15 °C
NS3 SEQUENCING FOR Primers	35 µL	Red	-25°C to - 15 °C
NS3 SEQUENCING REV Primers	35 µL	Red	-25°C to - 15 °C

Table 1: Volumes and storage conditions of the *DeepChek® Assay NS3 Genotyping and Drug Resistance V1 (RUO)*

## Reagent storage and handling

The **DeepChek® Assay NS3 Genotyping and Drug Resistance V1 (RUO)** is shipped with dry ice and should be maintained and stored immediately upon receipt at –20°C in order to avoid compromising cold chain integrity.

Expiration date: please refer to the label on the kit box.

## Materials required but not provided

- Thermocycler ABI9700
- 96-well plate cooler (Eppendorf / 22510509)
- 96-well PCR plates (Eppendorf / 951020303)
- Plates thermo seals (Thermo Scientific / AB-0558)
- Plate centrifuge
- 0.2 mL thin-wall 8 tubes & domed caps (Thermo Scientific / AB-0266)
- 1.5 mL centrifuge tubes (Dot Scientific Inc. / RN1700-GST)
- Centrifuges tubes (see your specific centrifuge manual)
- Mini centrifuge (see your specific centrifuge manual)
- Microliter pipets dedicated to PCR (0.1-2.5 µL; 1-10 or 1-20 µL; 20-200 µL)
- Ice

### Note:

- Ensure that instruments have been checked and calibrated according to the manufacturer's recommendations.
- Please refer to relevant manufacturer's Instructions for Use (IFU) to proceed with the instrument.

## Warnings and precautions

- **For Research Use Only (RUO). Not for use in diagnostic procedures.** No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of disease.
- Laboratories maybe required to report all test results to the appropriate public health authorities.
- Store assay reagents as indicated on their individual labels.
- Do not mix reagents from different kit lots.
- Reagents must be stored and handled as specified in these instructions for use. Do not use reagents past expiration date.
- Work surfaces and pipettes should be cleaned and decontaminated with cleaning products such as 10% bleach, "DNAzap™" or "RNase AWAY®" to minimize risk of nucleic acid contamination. Residual bleach should be removed using 70% ethanol.
- Use personal protective equipment (PPE) consistent with current guidelines for the handling of potentially infectious samples.
- Do not eat, drink, smoke, apply cosmetics or handle contact lenses in areas where reagents and human specimens are handled.
- Always use pipette tips with aerosol barriers. Tips that are used must be sterile and free from DNases and RNases.
- Dispose of waste in compliance with the local, state, and federal regulations.
- Frequent cleaning of the wells of the PCR instrument plate is recommended to prevent contamination.
- To avoid contamination, use separated and segregated working areas.

- Check whether the PCR reaction tubes are tightly closed before loading on the PCR instrument to prevent contamination of the instrument from leaking tubes.

## Starting

- Identify the product.
- Verify the expiration date.
- Verify the latest instruction for use available for the product lot number.
- Verify if the product was used already. If yes, check the remaining tests available.

## RNA Extraction

To achieve optimal and sensitive RNA analysis, the best representation of the viral quasispecies, it is recommended to extract **1 mL** of plasma for subsequent cDNA and amplicon generation and elute in the minimum volume required for your preferred extraction kit - MagNA Pure Compact Nucleic Acid Isolation Kit I (Roche Diagnostics) is recommended.

The **DeepChek® Assay NS3 Genotyping and Drug Resistance V1 (RUO)** will work with at least an extraction of 400µL of plasma or serum, ideally from fresh samples, to be eluted in 100µL (related sensitivity evaluated to 1250 UI/mL).


For samples with low viral load, we recommend:

- To perform an ultracentrifugation procedure. Pellet the virus for 1.5 hours at 40,000 g (or alternatively for 2 hours at 24,000g), and at 4°C. Remove enough supernatant to leave the required amount of sample for your preferred extraction kit.

**OR**

- To extract one or 2 mL of plasma/serum and elute in the minimum volume required for your preferred extraction kit.

## RT-PCR Workflow Overview



## RT and PCR Step-by-Step Workflow for NS3 Target

1. Thaw extracted template RNA, primer solutions, dNTP Mix, 5x Buffer and RNase-free water and place them on ice. Load all the tubes into the centrifuge. Spin the samples at 11000 g for 10 seconds. And then aspirate and discharge the solution several times before the dispensing.
2. Prepare a RT master mix according to **Table 2**. Prepare a volume of RT master mix greater (n+1) than that required for the total number of reactions to be performed.

Reagent	Volume
RT & PCR Buffer 2X	7.5 µL
RP Solution	0.3 µL
BP Solution	0.15 µL
RT Enzyme	0.3 µL
Water	1.75 µL

Table 2: Reaction components for RT-PCR NS3 target

3. Vortex the master mix thoroughly and dispense 10 µL into PCR tubes. Mix by pipetting the master mix up and down a few times.
4. Add 5 µL of RNA to the PCR tubes. Mix by pipetting the master mix up and down a few times.
5. Program the thermal cycler according to the program in **Table 3**.

Temperature (°C)	Time
30	5 min
42	5 min
95	15 sec
4	∞

Table 3: RT NS3 program

6. Start the RT NS3 program.
7. During RT NS3 reaction, prepare a PCR master mix according to **Table 4**. Prepare a volume of PCR master mix greater (n+1) than that required for the total number of reactions to be performed.

Reagent	Volume
RT & PCR Buffer 2X	7.5 µL
NS3 FOR1 PCR Primers	0.6 µL
NS3 FOR2 PCR Primers	0.6 µL
NS3 REV PCR Primers	1.2 µL
PCR Enzyme	0.3 µL
Water	4.8 µL

Table 4: Reaction components for RT-PCR NS3 target



8. Vortex the PCR master mix thoroughly and dispense 15 µL into RT product tubes. Mix by pipetting the master mix up and down a few times.
9. Program the thermal cycler according to the program in **Table 5**.

Cycle	Temperature (°C)	Time
1	94	3 min
45 cycles	94	15 sec
	53	1 min
	72	1 min
1	4	∞

Table 5: RT NS3 program

10. **[Recommended]** PCR products can be controlled through electrophoresis on an agarose gel. After amplification, samples can be stored overnight at 2–10°C, or at –20°C for long-term storage.

**Expected amplicons size: 776 bp.**

### Nested PCR Step-by-Step Workflow for NS3 Target

1. Prepare a NESTED PCR master mix according to **Table 6**. Prepare a volume of master mix greater (n+1) than that required for the total number of reactions to be performed.

Reagent	Volume
<b>Nested PCR Buffer 2X</b>	12.5 µL
<b>NS3 FOR1-NESTED-PCR Primers</b>	0.5 µL
<b>NS3 FOR2-NESTED-PCR Primers</b>	0.5 µL
<b>NS3 REV-NESTED-PCR Primers</b>	1 µL
<b>NESTED PCR Enzyme</b>	0.3 µL
<b>Water</b>	8.3 µL

Table 6: Reaction components for Nested PCR NS3 target

2. Vortex the NESTED PCR master mix thoroughly and dispense 23 µL into PCR tubes. Mix by pipetting the master mix up and down a few times.
3. Add 2 µL of RNA to the PCR tubes. Mix by pipetting the master mix up and down a few times.
4. Program the thermal cycler according to the program in **Table 7**.

Cycle	Temperature (°C)	Time
1	94	5 min
5 cycles	94	30 sec
	53	1 min
	72	1 min
40 cycles	94	15 sec
	67	30 sec
	72	30 sec
1	4	∞


Table 7: Nested PCR NS3 program

5. **[Recommended]** PCR products can be controlled through electrophoresis on an agarose gel. After amplification, samples can be stored overnight at 2–10°C, or at –20°C for long-term storage.

**Expected amplicons size: 650 bp.**

## RT-PCR Troubleshooting Guide

1. Check the concentration, storage conditions and quality of the starting template. For optimal results use fresh/frozen plasma/serum and proceed with fresh RNA extraction.
2. For samples with low viral load, we recommend to perform an ultracentrifugation procedure. Pellet the virus for 1.5 hours at 40,000 g and at 4°C. Remove enough supernatant to leave the required amount of sample for your preferred extraction kit.
3. In presence of very large PCR bands on the agarose gel, dilute ( $1/10^1 - 1/10^3$ ) of the product before sequencing.



## PCR Products Purification

Before sequencing, first make sure your Nested PCR products have been purified.

## Sequencing

### 1. Purification

For amplicons greater than 500bp, use a 0.6x AMPure XP cleanup (30 µl volume of beads). For amplicons less than 500bp, use a 1.8x AMPure XP cleanup (90 µl volume of beads) to maximize yield.

### 2. Sanger sequencing

In presence of very large PCR bands on the agarose gel, make dilution ( $1/10^1 - 1/10^3$ ) of the RT-PCR product before sequencing.

For NS3 sequencing, use the 2 red cap tubes containing 29µL of each NS3 Sanger sequencing primer (forward and reverse - 3.2 µM). **1 µL** of each primer will be used for each sample.

1. Prepare the sequencing reaction according to the **Table 8a** (Big Dyes Terminator kit v1.1) or **8b** (Big Dyes Terminator kit v3.1).

Reagent for <b>Forward</b> Sequencing	Volume
Big Dye Terminator v1.1	1 µL
Sequencing Buffer	1 µL
<b>Forward Primer (3,2µM)</b>	<b>1 µL</b>
Purified RT-PCR	0.7 – 2 µL
Water	q.s. to 10 µL

Reagent for <b>Reverse</b> Sequencing	Volume
Big Dye Terminator v1.1	1 µL
Sequencing Buffer	1 µL
<b>Forward Primer (3,2µM)</b>	<b>1 µL</b>
Purified RT-PCR	0.7 – 2 µL
Water	q.s. to 10 µL

Table 8a: Sequencing reaction for Big Dyes Terminator kit v1.1

Reagent for <b>Forward</b> Sequencing	Volume
Big Dye Terminator v3.1	2 µL
Sequencing Buffer	2 µL
<b>Forward Primer (3,2µM)</b>	<b>1 µL</b>
Purified RT-PCR	0.7 – 2 µL
Water	q.s. to 15 µL

Reagent for <b>Reverse</b> Sequencing	Volume
Big Dye Terminator v3.1	2 µL
Sequencing Buffer	2 µL
<b>Forward Primer (3,2µM)</b>	<b>1 µL</b>
Purified RT-PCR	0.7 – 2 µL
Water	q.s. to 15 µL

Table 8b: Sequencing reaction for Big Dyes Terminator kit v3.1

2. Program the thermal cycler according to the program in **Table 9a** (Big Dyes Terminator kit v1.1) or **9b** (Big Dyes Terminator kit v3.1).

Cycle	Temperature (°C)	Time
1	96	5 min
25 cycles	96	10 sec
	50	5 sec
	60	4 min
1	4	∞

Table 9a: Thermal cycler for Big Dyes Terminator kit v1.1

Cycle	Temperature (°C)	Time
1	96	1:30 min
25 cycles	96	20 sec
	50	15 sec
	60	1:30 min
1	4	∞

Table 9b: Thermal cycler for Big Dyes Terminator kit v3.1

3. **Sephadex purification** - Complete all the sequencing reaction with water (q.s. to 20µL). Purify all sequencing reaction (20µL) with Sephadex gel before the final Sanger sequencing.

**OR**

3. **PCR Cleanup reagent**

**OR**

3. **Ethanol purification**
  - a) Add 4ul of EDTA 125mM and sodium acetate 3M (1:1) solution;
  - b) Add 50ul Ethanol 100%;
  - c) Seal the plate well and gently vortex;
  - d) Incubate at room temperature (protected from light) for 15 minutes; While the plate is incubating, change the temperature of the centrifuge to 4°C;
  - e) Centrifuge the plate at 4°C for 35 minutes, at 4000rpm;
  - f) Remove the plate immediately, remove the sealing, and place the plate face down on a paper towel.
  - g) Briefly centrifuge the place face down until 185rpm;
  - h) Add 55ul of Ethanol 70%;
  - i) Seal the plate well and vortex for 15 seconds;
  - j) Centrifuge the plate at 4°C for 15 minutes, at 4000rpm;
  - k) Remove the plate immediately, remove the sealing, and place the plate face down on a paper towel;
  - l) Centrifuge the place face down for 1 minute at 4000rpm;
  - m) Place the plate at 94°C for maximum 1 minute.
  
4. **Denaturation:** Add 10 µL of formamide and incubate at the thermocycle at 94°C for 5 minutes then immediately incubate the plate ar 4°C for thermal shock for at least 5 minutes.

After these steps you can place the plate on the sequencer and start run.

### 3. NGS

After the amplicon verification, the samples are ready for the NGS kit processing:

Through Illumina MiSeq

- [116A24 / 116A48 / 116A96 | DeepChek® NGS LIBRARY PREPARATION V1 \(24 / 48 or 96 reactions\).](#)
- MS-103-1003 | MiSeq Reagent Nano kit, v2 (500 cycles).

Through Ion Torrent

- [4471269](#) | Ion Xpress™ Plus Fragment Library Kit
- [4471250](#) | Ion Xpress™ Barcode Adapters 1-16 Kit
- [4484355](#) | Ion 318™ Chip Kit v2

## Data Analysis

### 1. Sanger

AB1 or FASTA files containing nucleotide sequences for NS3 are analyzed by the **DeepChek®** software. User shall then follow the DeepChek software procedure to complete the data analysis and reporting processes.











## 2. NGS

NGS files containing nucleotide sequences for NS3 are analyzed by the **DeepChek®** software. User shall then follow the DeepChek software procedure to complete the data analysis and reporting processes.

### Product quality control

In accordance with ABL's Quality Management System, each lot of the assay is tested against predetermined specifications to ensure consistent product quality. Certificates of Analysis are available upon request.

### Symbols

	Contains reagents enough for <N> reactions		Consult instructions for use
	Caution		Temperature limitation
	Catalog number		Serial Number
	Use by	<b>Rn</b>	R is for revision of the Instructions for Use (IFU) and n is the revision number
	Manufacturer		Country of manufacturing
	Distributor		

### Contact Information

For technical assistance and more information, please see our Technical Support Center at Online: [https://ablsa.odoo.com/fr\\_FR/issue](https://ablsa.odoo.com/fr_FR/issue); Email: [support-diag@ablsa.com](mailto:support-diag@ablsa.com); Call +339 7017 0300 Or contact your ABL Field-Application Specialist or your local distributor. For up-to-date licensing information or product-specific disclaimers, see the respective ABL Assay User Guide. ABL User Guides are available at [www.ablsa.com/ifu](http://www.ablsa.com/ifu) or can be requested from ABL Technical Services or your local distributor.

### Manufacturer



Advanced Biological Laboratories (ABL) S.A.

17 rue des Jardiniers, L-1835 Luxembourg, Luxembourg

The customer is responsible for compliance with regulatory requirements that pertain to their procedures and uses of the instrument. The information in this guide is subject to change without notice. **DISCLAIMER: TO THE EXTENT ALLOWED BY LAW, ABL (S.A) AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.**

IFU\_108A24\_EN\_RUO © 2021 ABL S.A., all rights reserved.

Version 1.0

Effective date: 10<sup>th</sup> May 2021